Oxygen Concentrator

Operation Guide



Address: R 1 0 1 1 , X u e s o n g R o a d , High and New Tech,

Henan Xiaotaiyang Trading Co., Ltd

Zhengzhou City, Henan Province, China

Tel: 86 15838316518



Dear users,

The instructions introduce the function characteristics of this product, operation steps, notes and maintenance in detail.

In order to ensure that you can use the product more effectively, please read this manual carefully before use and keep it safe after reading.

Without permission, cannot take apart products in person, otherwise, you will lose the warranty.

Be careful, the company reserves the rights to update products and technology. Some illustrations in the specification you see in the product may be different, received in kind prevail.

Thank you for purchasing our products, wish you use happily.

Without reading this manual, please don't operate the product. If you have any questions on the understanding of this manual, please contact the seller or after-sales technical personnel, don't take action to operate the product, or it may result in serious personal injury or property damage.





Notes

Safety Precautions

▶ To prevent blackouts or household oxygen bar may break down, someone who are urgently needed oxygen must be configured with a much-needed backup support to other devices (such as: oxygen cylinder, oxygen bags, etc.), it is not considered for the life of the equipment or the use of the continuation of life.

▶ Oxygen may cause combustion, no smoking when operating the oxygen bar When in a heat environment, sparks and open flames and other factors, do

not use household oxygen bar.

There is high-voltage in household oxygen bar, please do not disassemble.

All parts of the product, after rigorous testing, safety and reliability, so do not use this product with other accessories made by other manufactures otherwise it will affect the performance.

Moving household oxygen bar by the cord or hose is prohibited.

This product with rated voltage of , \(\subseteq 220V, \subseteq 110V, \) please use secure socket qualified.


➤ To prevent accidents, do not let children operate the oxygen bar. When breathing oxygen, not let the hose around the neck. When the elderly using the oxygen, a guardian must beside him.

▶ Household oxygen bar works as a direct separation of oxygen from air, which is used for breathing, so do not use in smoke, toxic and harmful gases environments.

Cautions:

- ▶ Before the first use, please roll household oxygen bar, then cut off the bottom of the transport fixed ties, see "Cut ties description".
- ▶ When this product is working, there will be intermittent exhaust sound, which is a normal phenomenon.
- ▶ When add water to wet cup, do not exceed the maximum water level. When handling household oxygen bar, need to pour out the water in wet cup, so as not to cause backflow.
- ▶ Please confirm whether the oxygen outlet, the air intake vents behind oxygen bar is blocked, in order to avoid downtime caused by high temperature.
- ► This product only applies to household oxygenating, cannot be used as medical treatment.

Catalogue





Product Features

Product Features

Product Introduction

This product uses the air as raw material, $AC \square 220V \square 110V$ power as power source, uses a good original zeolite by French and the air compressor 5000 hours lifespan, Through advanced pressure swing adsorption separation in room temperature, extraction of high-purity oxygen by air, stable oxygen, safe and reliable, this method is called pure physical methods. (Refers to working principle on page 12).

Functional characteristics

Intelligent Control:

IMD beautiful large color panel, easy to operate, large color LED screen, clear display, timer operation function and with infrared remote control, let it be easier and more convenient.

Anion:

This machine is equipped with ion function, and a "negative" button; the neg ative ion system can work alone , you can also work simultaneously with the oxygen system; Anion generator Air vents located in the machine, the exhaus t vents discharged into the surrounding space of the machine when working.

Multi-layer filter, self-easy to replace:

The oxygen system of this product has coarse dust filter, fine dust filter and three bacterial filtration treatment for the input air respectively, finally, the oxygen is fresh and clean after filtering, and the two front layers filter can be replaced without disassembing, the user operate it comfortably.

Using environment

Temperature: 5-35℃ Relative humidity: ≤80% Range of the atmospheric pressure: 0.086 MPa -0.106 MPa When the temperature is higher or lower, do not use oxygen bar.

Technical parameters

Product name: household Oxygen Bar

Model: □MAF015A □MAF015AW

□ MAF015BW □ MAF015BJ

Scope of application: use as household oxygenating,

not use for medical treatment

PSA methods: PSA
The input power: 115W

Rated voltage: AC 220 v, 50 Hz□ AC 110V, 50Hz□

Current rating: 0.5 A
Air output: 1~6L/min
Output pressure: <100kpa

Negative oxygen ion: ≥ 6 million particles/ cm³

Noise: about 48dB

Product size: 302 * 168 * 325 mm

Net weight: 5.2kg Electrical categories: II

Execution standard: GB4706.1-2005

Q/MAF 01-2017

Functional difference showed in the following table

Model	MAF015A	MAF015AW	MAF015BW	MAF015BJ		
Oxygen concentration	30%~90%	30%~90%	30%~90%	30%~90%		
Remote control function	Yes	Yes	Yes	Yes		
Spray function	No	Yes	Yes	Yes		
Cumulative function	Yes	Yes	Yes	Yes		


Installation and Operation


Installation and Operation

Opening instructions

Oxygen packing, please check the items in the box whether they complete or in tact according to the manual attachment list. Otherwise, please contact retailer or manufacturer.

Appearance diagram:





-04- English -05-

Installation and Operation


Power operation and components using

► Wet cup (add water):

Use left hand to hold wet cup, use right-hand to grip and rotate wet cup counterclockwise, lid open after the rotation in place, adding moderate amount of pure water, then close the lid and turn clockwise rotation in place.



Use of wet cup;



Note: "The machine connected to the outlet of which ends with the wet cup white tube"

Installation and Operation

Put one end of the conduit to the wet cup inlet port B (with the inlet conduit, as difference) A is inserted into the other end of the conduit connected to the oxygen outlet of household oxygen bar, and finally the end of the oxygen duct plug to the integrated water heater outlet E.


When the water level inside the wet cup exceeds the maximum water level D, please promptly drain the water to avoid water back into the oxygen tube.

► Connect the power cord, power on

Put the tail end of the power cord into the power cord socket household oxygen bar, and the other end connected to the indoor outlet, after power, the power switch is turned on, after the first power, microcomputer will automatically start the self-test all the bright lights of a display about 2 seconds, colleagues buzzer will sound prompt, after the self-test, automatically restore the standby mode and displays the default time "015", "power jack" and "power switch" position, see section appearance of page four.

▶ Power on: ON/OFF button

▶ Time



In standby mode, press the "ON / OFF" to wake up the machine and start working, and "oxygen" light will be lit, the default length of 15 minutes, turn off the machine by complex back to standby.

(月间) or

Press "+, -" button to adjust the time of work time, after finish setting, press the "On / Off" button, the oxygen machine starts to work. (after the setup is completed, wait three seconds, the oxygen machine will automatically start working) Click the "timing" until showing the length of time. After finish working , the display window will return to standby mode and waits for the user to use: If you do not use for more than ten minutes (temperament cooling system is running for about five minutes), it will enter power-saving sleep state (anion is closed), then "Power" will flash to alert the user who can press the power switch back to the off state.

-06- English -07-

Installation and Operation







or



According to users' needs, you can open anion function, click "anion" to open, and display "anion" light will be lit (see screen instructions on page five), press twice means close.

Remote control use



Infrared receiver



With a dedicated remote control, this product gets the remote control alignment infrared receiver hole of oxygen machine panel.

When you find the remote control operation insensitive and cannot operate, you must check the remote control battery whether it fully charged first; remote control can't be too far from the host computer, otherwise it will lead to the remote control operation insensitive.

Routine maintenance

- ◆ The oxygen machine has smooth appearance, glossy, please prevent the sharp scratch it.
- Please turn off the oxygen machine after finishing the timer and the air supply, and unplug the power in time.
- The cotton filters on the host side change in time, if the cotton filters were blocked by the dirty stuff, it will have an effect on oxygen.

Attention:

Please don't take apart the machine by yourself, avoid unnecessary damage inside the machine, before cleaning and maintaining, oxygen machine must be cut off the power.

Handle of long idle

- Remove the wet cup, drain the water out of the cup and replace host computer after air-dry.
- Starting up, let the oxygen machine work for thirty minutes, then wait for it to cool by natural.
- Cut off the power, replace the oxygen machine into the box.
- Remote control can't install battery while sealing, in order to prevent the battery leakage to damage the remote control.
- Pls wrap and seal the main engine with plastic bags to prevent air humidity cause the machine performance degradation.

Parts maintenance

- ◆ The maintenance of wet cup:
 In order to ensure the cleaning of the wet cup, the injection should be pure water or boiled water, and try to replace every day, we suggest users cleaning the wet cup once a week
- Cotton filters and filter maintenance:
- The main function of the filter is filter and adsorption the dust and impurity, it will form a blockage in turn to affect the intake of oxygen machine for a long time.

It will become intake block for a long time, thereby, have an effect on the oxygen machine; generally, replaced once a quarter, if it has much dust, recommend replacing every two months.

-08- English -09-

Problem Investigation and Trouble Shooting


Problem Investigation and Trouble Shooting

Symptom	Possible Cause	Solution				
Household oxygen	The power line is not connected to a power outlet.	Check whether the socket outlet connection is normal.				
bar cannot operate	Power supply socket without electricity.	Replace the electric socket.				
properly, the power indicator light is not	Overload protection burning fuse	Open the insurance and replace the fuse.				
bright.	Power cord is not for good in the household oxygen bar or power cord bad.	Unplug reinsert the fuselage outlet or replace a power supply cord.				
The screen display normal, but the oxy- gen bar does not ha- ve normal operation	Household oxygen bar over- heating protection.	Make sure the place has been away from ho- usehold oxygen bar curtains, bake warm m- achines, heaters and fireplaces, and confirm household oxygen bar and walls or other o- bjects around the retention gap of at least 20 cm.				
soun.	Household oxygen bar com- puter froze plate.	Turn off the machine, wait five minutes to restart.				
	The compressor function b- reakdown.	Contact household oxygen bar retailer or manufacturers.				
Full screen flash, while alarm.	Internal hardware damage	Returned to the manufacturer or dealer for repair.				
Big Noise	Ties are not cut, the machine did not put a smooth.	Cut ties, put the machine steady.				
The screen display normal and hear n- ormal operation so-	Intake window cotton filters is dirty, cause block.	Household oxygen bar internal malfunction.				
und but no anoin o- xygen output.	Cut off the power switch, replace window air intake filter cotton.	Contact the household seller or manufacturers.				
Press to adjust the time when the tra- nsfer is invalid.	Sleep tone when keys are locked, inoperable.	Press the "On / Off" button to wake, the key to adjust the time and then tune.				

Symptom	Possible Cause	Solution					
The oxygen machine emits poof bared sound six-minute intervals.	The machine emits exhaust sound.	If the machine emits poof bared sound reg- ularly is a normal phenomenon, discharge the sound of nitrogen.					
Water coming out of the mouth of o- xygen.	Adding too much water in the wet cup (exceed the highest water level),airflow will push oxygen tube blisters.	The water in the wet cup have reduced to the highest water level.					
	Difference temperature , ca- use the humidified oxygen in the wall that moisture will c- ondense water out of cold.	Pour the water out of the wet cup, starting up, after fifteen minutes, waiting the droplets are dried in the oxygen tube, then adding water into the wet cup.					
the machine is op-	Flow value is too small.	Turn up flow					
erating normally - after starting up, but have a narrow minded.	Wet cup is not installed correctly cause leak	wet cup cover is not fit in place, unscrew the lid anticlockwise, after checking the right direction again wet glass bottles. wet cup body sets a ring, if you install the missing equipment, it can cause leak.					
Oxygen odor	Early use, series of tubes or other plastic will emit odor, is a normal phenomenon.	This odor is normal smell emitted by the silicone tube, which has been detected by non toxic, and with the use of the machine, the smell will gradually disappear.					

-10- English -11-

Working Principle



Warranty Card

Dear user:

Thank you for purchasing this product, for the future use of this product, you will enjoy the company to provide you with high quality service, in order to use the product better, please be sure to carefully read the "warranties" before your use.

		$\langle \rangle \langle \rangle$				Puro	chase	Date	e:		year		nont	h	day	
\ \ \	User Name						Sex									
0	Age					\times	ontac epho	$ \times$	$\stackrel{\sim}{\sim}$							> >
000	Mobile Phone						Zip									
0	Contact Address								2							
\ \ \	Product Model						tem N	10								
	Sales De-	XX				S	eller		\Rightarrow		$\langle \rangle$		$\langle \rangle$	$\langle \rangle$		
	partment	X~X	<u>~~~</u>	<u> </u>	<u> </u>	P	hone		\sim	\geq		$\geq \langle$	\sim	<u> </u>		
0	Notes:															+
$^{\circ}$	XXX															
\sim	XUXUX.															6
_			~X~?			$\langle \rangle \langle \rangle$			\sim			$> \langle$		> <		6

Users save

	Purchase Date:	year	month	day
User Name	Sex			
Age	Contact Telephone			
Mobile Phone	Zip	8-8		8-8
Contact		2-2		2-2
Address		XX		$\times \times$
Product Model	Item No			
Sales De-	Seller	$\Diamond \circ \Diamond \circ$		$\Diamond \neg \Diamond$
partment	Phone	2 - 2	2000	\mathcal{Y}
Notes:				

Businesses save

Warranty Issues

- 1. Warranty: main engine with one year warranty.
- 2. If you buy our machine, you can enjoy the whole year warranty provided by the company's policy.
- 3. You can register information via the internet or by mail, for warranty.
- If the items listed below, based on the cost of fees and subject to close for repair.
- 1. out of warranty period of the product;
- 2. failure to use the product manual and product damage.
- Under one of the following situations, Not within the warranty service.
- 1. unauthorized alteration warranty certificate, serial number and product labels;
- 2. damaged products caused by unauthorized hardware changes (items missing, variation, etc);
- 3. due to damage caused by external products;
- 4. due to force majeure (earthquake, fire, flood, war and other disasters) product damage;
- 5. Unauthorized disassemble products;
- 6. deliberate destruction caused by human fault products;
- 7. The failure to require the installation and use instructions damage caused by the company's products;

Certificate

Production Date:


Quality Inspector:

Version Number:1.1 Part No:00.04.04-00605-02

Cut Ties Description

Before using it for the first time, please follow the steps to cut cable ties.






1. The arrow points to tie for transport.



2. Use pliers to cut fixed ties.



3. Ties must be drawn out.



4. Two fixed ties can be used after being extracted.

Serial No:

Parts No. 428-00503-000E