LASER TREE



525nm/5W/Fiber Coupled Lasers /With TEC/Unpluggable Fiber



Features

- 525nm Green Diode Lasers
- Built-in TEC Cooling
- 60um,100um,200um,400um Selectable With Unpluggable Fiber

Applications

- Bio & Medical
- Special Light Source
- Light Source Of Optical Equipments

Absolute Maximum Ratings

Parameter	Symbol	Value	Unit
Reverse Current	I _r (LD)	85	mA
Operating Temperature	T _{opr}	0 ~ +65	°C
Storage Temperature	T _{stg}	-40 ~ +85	°C
Lead soldering temperature (10 sec.)	T _{sld}	260	°C

Data sheet Rev1.0 Apr.2023

Tel: +86 136 8264 1046

E-mail: lasertree@micost-optotech.com



■ Optical and Electrical Characteristics (Tc=25[°]C)


Parameter	Symbol	Min	Тур	Мах	Unit	Test Condition
Center Wavelength	λς	520	525	530	nm	Po=5W
Spectral Width (FWHM)	Δλ		2.0	_	nm	Po=5W
Optical Output Power	Po	4.5	5.0	5.5	w	lop=2.3A
Fiber Type	_	N	1ulti Mode Fibe	er	_	_
Fiber Core Diameter	_	60,100	,200,400 se	lectable	um	_
Numerical Aperture	NA		0.22		_	_
Fiber Connector	_	SMAS	905,FC selec	table		_
Working Condition			CW			_
Thermistor	Rt		50		К	_
Threshold Current	lth	150		500	mA	_
Operating Current	Іор		1.9	2.3	A	Po=5W
Operating Voltage	Vop	16		24	V	Po=5W
TEC Current Max	lop(TEC)		_	5.6	A	_
TEC Operating Max	Vop(TEC)			30.8	V	_


Data sheet Rev1.0 Apr.2023

Tel: +86 136 8264 1046 E-mail: lasertree@micost-optotech.com



Outline Dimensions





Data sheet Rev1.0 Apr.2023

Tel: +86 136 8264 1046 E-mail: lasertree@micost-optotech.com

Cautions

- Laser light can damage the human eye and skin. Do not expose the eye and skin to any laser light directly or through optical lenses. Focused laser beam through optical instruments will increase the chance of eye hazard.
- 2. The laser needs a stable driver to avoid surges. The instantaneous reverse current and reverse voltage should not exceed the absolute maximum ratings, otherwise it will damage the laser diode.
- 3. Semiconductor lasers are sensitive to temperature, working at high temperatures will reduce the photoelectric conversion efficiency and accelerate the aging of laser diode. It should be used under adequate heat dissipation or cooling conditions.
- **4.** The laser should be used at the rated current and rated power, the output power is too high will accelerate the aging of laser diode.
- **5.** Semiconductor lasers are electrostatic sensitive devices, anti-static measures should be taken during storage and transportation.
- 6. Lasers should be stored or work in a dry, ventilated environment to prevent damage to the laser from condensation.
- 7. The light emitting surface (cavity surface) is the key parts of the laser diode, any handling to damage the cavity surface should be strictly prohibited. It should be ensured that the laser diode is not contaminated, and damaged in mechanical.
- **8.** The fiber should not be bent at large angles, it should be ensured that the diameter of the bend should be more than 300 times the diameter of the fiber.



Data sheet Rev1.0 Apr.2023

Tel: +86 136 8264 1046 E-mail: lasertree@micost-optotech.com